


## Ordering code BAROLI 05

31.5.2016

### BAROLI 05

Battery powered digital pressure gauge (0 ... 0,4 / 600 bar)



Pressure									
gauge	M	0	0						
absolute	M	0	1						
Input [bar]									
0 ... 0,4		4	0	0	0				
0 ... 0,6		6	0	0	0				
0 ... 1		1	0	0	1				
0 ... 1,6		1	6	0	1				
0 ... 2,5		2	5	0	1				
0 ... 4		4	0	0	1				
0 ... 6		6	0	0	1				
0 ... 10		1	0	0	2				
0 ... 16		1	6	0	2				
0 ... 25		2	5	0	2				
0 ... 40		4	0	0	2				
0 ... 60		6	0	0	2				
0 ... 100		1	0	0	3				
0 ... 160		1	6	0	3				
0 ... 250		2	5	0	3				
0 ... 400		4	0	0	3				
0 ... 600		6	0	0	3				
-1 ... 0		X	1	0	2				
Customer		9	9	9	9				
Accuracy									
0,50%									5
0,5% calibration including certificate									T
1% (PN < 0 bar)									8
Customer									9
Mechanical connection									
G 1/2" EN 837							2	0	0
G 1/4" EN 837							4	0	0
M 20x1,5 EN 837							8	0	0
1/2" NPT							N	0	0
1/4" NPT							N	4	0
Customer							9	9	9
Seals									
Viton (FKM) (PN < 100 bar)									1
NBR (PN ≥ 100 bar)									5
Customer									9
Pressure port									
stainless steel 1.4404 (316 L)									1
Customer									9
Diaphragm									
Ceramic Al2O3 96%									2
Customer									9
Front foil									
Standard (design BD Sensors)									1
Neutral									N
Special version									
Standard									0
Customer									9
Accessories									
Battery 2pcs									1000377

0,-...without additional charge  
 On request...in accordance with the producer  
 Additional charge for calibration is not subject to reduction.  
 This data sheet contains product specification; properties are not guaranteed.  
 Subject to change without notice.



BD SENSORS s.r.o.  
 Hradištská 817  
 CZ – 687 08 Buchlovice

Tel.: +420 572 411 011  
 Fax: +420 572 411 497

www.bdsensors.cz  
 info@bdsensors.cz

The company BD SENSORS s.r.o. is certified by TÜV SÜD Czech according to the standard ISO 9001.

