

Ordering code DM 01-500

3.1.2020

DM 01-500 digital display for DM 01

Precision pressure transmitter (0 ... 0,1 / 400 bar)									
Pressure									
Gauge	M	0	K						
Absolute (PN ≥ 0,4 bar)	M	0	L						
Input [bar]									
0 ... 0,1				1	0	0	0		
0 ... 0,16				1	6	0	0		
0 ... 0,25				2	5	0	0		
0 ... 0,4				4	0	0	0		
0 ... 0,6				6	0	0	0		
0 ... 1				1	0	0	1		
0 ... 1,6				1	6	0	1		
0 ... 2,5				2	5	0	1		
0 ... 4				4	0	0	1		
0 ... 6				6	0	0	1		
0 ... 10				1	0	0	2		
0 ... 16				1	6	0	2		
0 ... 25				2	5	0	2		
0 ... 40				4	0	0	2		
0 ... 60				6	0	0	2		
0 ... 100				1	0	0	3		
0 ... 160				1	6	0	3		
0 ... 250				2	5	0	3		
0 ... 400				4	0	0	3		
-1 ... 0				X	1	0	2		
Customer				9	9	9	9		
Version									
Standard									0
Intrinsic safety Ex ia									E
Accuracy									
0,05% BFSL (PN ≥ 0,4 bar)									B 1
0,125% BFSL (PN < 0,4 bar]									B 2
Customer									9 9
Mechanical connection									
G 1/2" DIN 3852									1 0 0
G 1/2" EN 837-1/-3 (manometric)									2 0 0
G 1/4" DIN 3852									3 0 0
G 1/4" EN 837-1/-3 (manometric)									4 0 0
G1/2" DIN 3852 with flush sensor (PN ≥ 0,4 bar)									F 0 0
G 1/2" DIN 3852 open pressure port									H 0 0
1/2" NPT									N 0 0
1/4" NPT									N 4 0
G 1/4" EN837 internal thread, welded (PN ≥ 0,4 bar)									J 0 3
Customer									9 9 9
Seals									
Viton (FKM)									1
Customer									9
Special version									
Standard									5 0 0
Cutomer									9 9 9

0,-... without additional charge

On request...in accordance with the producer

Additional charge for calibration is not subject to reduction.

This data sheet contains product specification; properties are not guaranteed.

Subject to change without notice.



BD SENSORS s.r.o.
Hradištská 817
CZ – 687 08 Buchlovice

Tel.: +420 572 411 011
Fax: +420 572 411 497

www.bdsensors.cz
info@bdsensors.cz



The company BD SENSORS s.r.o. is certified by TÜV SÜD Czech according to the standard ISO 9001.