

Ordering code LMP 808

31.5.2016

LMP 808

Submersible transmitter made out of PVC (0 ... 0,1 / 10 bar)



Pressure

gauge
measured value in m H₂O

4	1	0
4	1	1

Input	[mH ₂ O]	[bar]	[][][]	[][][]	[]	[]	[]	[][][]	[][]
	1.0	0.10	1	0	0	0			
	1.6	0.16	1	6	0	0			
	2.5	0.25	2	5	0	0			
	4.0	0.40	4	0	0	0			
	6.0	0.60	6	0	0	0			
	10	1.0	1	0	0	1			
	16	1.6	1	6	0	1			
	25	2.5	2	5	0	1			
	40	4.0	4	0	0	1			
	60	6.0	6	0	0	1			
	100	10	1	0	0	2			
	Customer		9	9	9	9			

Housing

PVC grey A

Diaphragm

Stainless steel 1.4435 1

Output

4 ... 20 mA / 2 wire	1
0 ... 20 mA / 3 wire	2
0 ... 10 V / 3 wire	3
Customer	9

Seals

Viton (FKM)	1
EPDM	3
Customer	9

Electrical connection

Without cable part	0
PVC - cable (price for 1 m)	1
PUR - cable (price for 1 m)	2
FEP - cable with teflon sheath (price for 1 m)	3
Customer	9

Accuracy

0,50% (P _N ≤ 0,4 bar)	5
0,35% (standard) (P _N > 0,4 bar)	3
by agreement 0,25% (P _N > 0,4 bar)	2
0,50% calibration including certificate (P _N ≤ 0,4 bar)	T
0,35% calibration including certificate (P _N > 0,4 bar)	S
by agreement 0,25% calibration including certificate (P _N > 0,4 bar)	R
	N
	M

Measured values table for accuracy 0,35% (only on customer request)

Customer 9

Cable length

without cable	0	0	0
in whole meters, e.g. 3 m	0	0	3
in whole meters, e.g. 15 m	0	1	5

Special version

Standard	0	0	0
Customer	1	0	6
	9	9	9

Accessories for submersible transmitter

Cable part + price for cable in whole meters	5000695
	5000992
Terminal clamp - zinc plated	1003440
Terminal clamp - stainless steel 1.4301	1000278

0,-...without additional charge On request...in accordance with the producer
Additional charge for calibration is not subject to reduction. Subject to change without notice.
This data sheet contains product specification; properties are not guaranteed.

This document contains product specification; properties are not guaranteed. Detailed information about options are defined in the datasheet. Subject to change without notice.



BD SENSORS s.r.o.
Hradištská 817
CZ – 687 08 Buchovice

Tel.: +420 572 411 011
Fax: +420 572 411 497

www.bdsensors.cz
info@bdsensors.cz

The company BD SENSORS s.r.o. is certified by TÜV SÜD Czech according to the standard ISO 9001.

