

[1] **EC-TYPE EXAMINATION CERTIFICATE**  
according to Directive 94/9/EC, Annex III



(Translation)

[2] Equipment and Protective Systems intended for use in Potentially Explosive Atmospheres, Directive 94/9/EC

[3] EC-Type Examination Certificate Number: **IBExU10ATEX1122 X**

[4] Equipment: **Pressure transmitter**  
Type DX9-DMP..., DX9-DMK..., DX9-LMP..., DX9-LMK...and  
DX9- 17.600G

[5] Manufacturer: **BD SENSORS s.r.o.**

[6] Address: **Hradistska 817  
687 08 Buchlovice  
CZECH REPUBLIC**

[7] The design of the equipment mentioned under [4] and any acceptable variation thereto are specified in the schedule to this EC-Type Examination Certificate.

[8] IBExU Institut für Sicherheitstechnik GmbH, NOTIFIED BODY number 0637 in accordance with article 9 of the Council Directive 94/9/EC of 23 March 1994, certifies that the under [4] mentioned equipment has been found to comply with the Essential Health and Safety Requirements relating to the design and construction of equipment intended for use in potentially explosive atmospheres given in Annex II to the Directive.  
The test results are recorded in the test report IB-10-3-065 of 31 August 2010.

[9] Compliance with the Essential Health and Safety Requirements has been assured by compliance with EN 60079-0:2009, EN 60079-11:2007 and EN 60079-26:2007.

[10] If the sign "X" is placed after the certificate number, it indicates that the equipment is subject to special conditions for safe use specified under [17] in the schedule to this EC-Type Examination Certificate.

[11] This EC-Type Examination Certificate relates only to the design and construction of the specified equipment. If applicable, further requirements of this directive apply to the manufacture and supply of this equipment.

[12] The marking of the equipment mentioned in [4] shall include the following:

All types, except DX9-DMK KRO: **II 1G Ex ia IIC T4 Ga**

Type DX9-DMK KRO: **II 1G Ex ia IIB T4 Ga**

**IBExU** Institut für Sicherheitstechnik GmbH  
Fuchsmühlenweg 7 - 09599 Freiberg, Germany  
☎ +49 (0) 3731 3805-0 - 📠 +49 (0) 3731 23650

Authorised for certifications Explosion protection

By order

(Dr. Wagner)

**Schedule**



Freiberg, 31 August 2010

Certificates without signature and seal are not valid.  
Certificates may only be duplicated completely and unchanged. In case of dispute, the German text shall prevail.

[13]

**Schedule**

[14]

**to the EC-TYPE EXAMINATION CERTIFICATE IBExU10ATEX1122 X**


[15]

**Description of equipment**

The Pressure transmitters DX9 are different pressure measurement devices or level sonde in stainless steel enclosure with different pressure ports. As measuring cell and evaluation electronic serves a separately approved electronic modules. The devices are intended for use in potentially hazardous areas, where Category 1G devices are required. They are supplied by an intrinsically safe power supply of the Category „ia“.

Type code:

DX9- D M P nnn I



- D - pressure measurement devices; L – level sonde
- M - transmitter
- P – stainless steel cell; K - ceramic cell
- nnn - design
- I - precision; P - process pressure port

Exception:

DMK457 KRO with EMV 45 as clamp platine  
 17.600G conform to DMP

Type extent:

Name	Type	Electronic-module	Measurement cell
DMP 343	DMU	ELMO-STA	DSP 210 / ELV 66 = DSP 210 Z
DMP 331	DMU	ELMO-STA	DSP 4XX / ELV66 = DSP 4XX Z alternative ICS 86 / ELV 70 = ICS 86 Z
DMP 333	DMU	ELMO-STA	DSP 4XX / ELV 66 = DSP 4XX Z
DMP 335	DMU	ELMO-STA	MSP 12X / ELV 57 = MSP 12X Z
17.600G	DMU	ELMO-STA	MSP 12X / ELV 57 = MSP 12X Z
DMK 331	DMU	ELMO-STA	DSK 511 / ELV 58 = DSK 511 Z
LMP 331	DMU	ELMO-STA	DSP 4XX / ELV66 = DSP 4XX Z alternative ICS 86 / ELV 70 = ICS 86 Z
LMK 331	PS	ELMO-STA	DSK 511 / ELV 58 = DSK 511 Z
LMK 307	PS	ELMO-STA	DSK 511 / ELV 58 = DSK 511 Z
DMP 331 I	DMU	ELMO-PRE	DSP 4XX / ELI 20 = DSP 4XX X alternative ICS 86 / ELI 20 with aid platine PD16
DMP 333 I	DMU	ELMO-PRE	DSP 4XX / ELI 20 = DSP 4XX X
LMP 331 I	PS	ELMO-PRE	DSP 4XX / ELI 20 = DSP 4XX X alternative ICS 86 / ELI 20 with aid platine PD16
DMP 331 P	DMU	ELMO-STA	DSP 4XX / ELV66 = DSP 4XX Z alternative DSP 1XX / ELV66 = DSP 1XX Z
DMK 331 P	DMU	ELMO-STA	DSK 511 / ELV 58 = DSK 511 Z
DMK457 KRO	DMU	ELMO-STA	DSK 511 / ELV 58 = DSK 511 Z additional with EMV 45 as clamp platine.
LMP 307	PS	ELMO-STA	DSP 4XX / ELV66 = DSP 4XX Z alternative ICS 86 / ELV 70 = ICS 86 Z
LMP 307 I	PS	ELMO-PRE	DSP 4XX / ELI 20 = DSP 4XX X alternative ICS 86 / ELI 20 with aid platine PD16

**IBExU Institut für Sicherheitstechnik GmbH**  
An-Institut der TU Bergakademie Freiberg

LMP 308	PS	ELMO-STA	DSP 4XX / ELV66 = DSP 4XX Z alternative ICS 86 / ELV 70 = ICS 86 Z
LMP 308 I	PS	ELMO-PRE	DSP 4XX / ELI 20 = DSP 4XX X alternative ICS 86 / ELI 20 with aid platine PD16

**Technical Data**

supply electric circuit in type of protection Intrinsic Safety Ex ia IIC  
(supply + and -)

$U_i$  28 V DC

$I_i$  93 mA

$P_i$  660 mW

effective inner capacity

$C_i$  negligible

effective inner inductivity

$L_i$  negligible

plus line inductivities 1  $\mu$ H/m and line capacities 160 pF/m (cable supplied by the manufacturer)

The supply connections have an inner capacity of max. 27 nF to the housing.

Type DX9-DMK457 KRO in type of protection Intrinsic Safety Ex ia IIB

effective inner capacity

$C_i$  105 nF

effective inner inductivity

$L_i$  5  $\mu$ H

The supply connections have an inner capacity of max. 91 nF to the housing.

Ambient temperature range -20 °C to +70 °C (type DX9-\*\*\* I to +65 °C)

[16] **Test report**

The test results are explained in detail in the test report IB-10-3-065. The test documents are part of the test report and listed there.

Summary of the test results:

The highest-pressure transmitters DX9 fulfil the requirements of type of protection Intrinsic Safety ,ia' for an electrical equipment of the Equipment Group II, Category 1G, Explosion Group IIC or IIB and Temperature Class T4.

[17] **Special conditions**

- The equipment designed with connector have to be installed in such a way that the degree of protection IP 20 is always kept.
- The safety and assembly instructions contained in the operating instruction and the ambient temperature range  $-20\text{ °C} \leq T_a \leq +70\text{ °C}$  or at the types DX9-\*\*\* I  $-20\text{ °C} \leq T_a \leq +65\text{ °C}$  have to be taken into account.
- The device may be operated in explosive atmospheres which requires equipment of Category 1 only when there are atmospheric conditions (temperature of  $-20\text{ °C}$  to  $+60\text{ °C}$ , pressure of 0.8 bar to 1.1 bar).

[18] **Essential Health and Safety Requirements**

Confirmed by compliance with standards (see [9]).

By order

Freiberg, 31 August 2010



(Dr. Wagner)

[1] **1<sup>st</sup> Addition to  
EC-TYPE EXAMINATION CERTIFICATE IBExU10ATEX1122 X  
- Translation -**



[2] **Equipment:** **Pressure transmitter**  
Type DX9-DMP..., DX9-DMK..., DX9-LMP..., DX9-LMK...and  
DX9-17.600G

[3] **Manufacturer:** BD SENSORS s.r.o.

[4] **Address:** Hradistska 817  
687 08 Buchlovice  
CZECH REPUBLIC

[5] **Addition / Alteration**

The additions / alterations of the equipment stated under [2] compared to the equipment already certified comprise in particular:

- a) Use in the presence of combustible dust.
- b) Addition of types DX9-DMP457 and DX9-DMK457 under use a alterations EMC-board.  
The table are following added:

Name	Type	Electronic-module	Measurement cell
DMP 457	DMU	ELMO-STA	DSP 4XX / ELV66 = DSP 4XX Z alternative ICS 86 / ELV 70 = ICS 86 Z EMV 58 replaced by EMV 26 of ELMO-STA
DMK 457	DMU	ELMO-STA	DSK 511 / ELV 58 = DSK 511 Z EMV 58 replaced by EMV 26 of ELMO-STA

The effective inner capacity increases with these equipment to C<sub>i</sub> 105 nF. The supply connections have an inner capacity of max. 140 nF opposite the enclosure.

- c) Compact field enclosure as new variant of enclosure

[6] **Test Report**

The proof of explosion protection of the equipment stated under [2] is documented in the Test Report IB-10-3-349 of 8 March 2011. The test documents are part of the Test Report.

[7] **Test result**

IBExU certifies, that the equipment stated under [2] fulfill the Essential Health and Safety Requirements given in Annex II to the Directive 94/9/EC by compliance with EN 60079-0:2009, EN 60079-11:2007 and EN 61241-11:2006.

The equipment stated under [2] fulfil the requirements of type of protection Intrinsic Safety ,ia' for an electrical equipment of the Equipment Group II, Category 1G, Explosion Group IIC or IIB and Temperature Class T4 as we as Category 1D with a maximum surface temperature of 85 °C.

The marking of the equipment stated under [2] shall include the following:

All types, except DX9-\*457\*:  
 II 1G Ex ia IIC T4  
 II 1D Ex iaD 20 T 85 °C

Type DX9-\*457\*:  
 II 1G Ex ia IIB T4  
 II 1D Ex iaD 20 T 85 °C

**IBExU Institut für Sicherheitstechnik GmbH**  
An-Institut der TU Bergakademie Freiberg

This addition is only valid in combination with the EC-Type Examination Certificate  
IBExU10ATEX1122 X of 31 August 2010.

IBExU Institut für Sicherheitstechnik GmbH  
Fuchsmühlenweg 7 - 09599 Freiberg, Germany  
☎ +49 (0) 3731 3805-0 - 📠 +49 (0) 3731 23650

Authorized for certifications  
- Explosion protection -

By order



(Dr. Wagner)



(Identification No. 0637)

Freiberg, 8 March 2011

Certificates without signature and seal aren't valid.  
Certificates may only be duplicated completely and unchanged.  
In case of dispute, the German text shall prevail.

[1] **2<sup>nd</sup> Addition to  
EC-TYPE EXAMINATION CERTIFICATE IBExU10ATEX1122 X  
- Translation -**



[2] Equipment: **Pressure transmitter**

a) Types corresponding IBExU10ATEX1122 X: DX9-DMP..., DX9-DMK..., DX9-LMP...,  
DX9-LMK...and DX9-17.600G

b) Added/appended types: DX9-DMP 339, DX9-17.605 and DX9-26.600

[3] Manufacturer: **BD SENSORS s.r.o.**

[4] Address: **Hradistska 817  
687 08 Buchlovice  
CZECH REPUBLIC**

[5] **Addition / Alteration**

The additions / alterations of the equipment stated under [2] b) compared to the equipment already certified comprise in particular:

Addition of types DX9-DMP 339, DX9-17.605 and DX9-26.600 under use of new measuring cells as well as electronic configurations on the basis of the partial certification IBExU10ATEX1026 U incl. 1<sup>st</sup> Addition.

The table is added as follows:

Name	Type	Electronic-module	Measuring cell
DMP 339	DMU	ELMO-STA	ICS 89 / ELV 66 / ELV 68 / EMV 26
		ELMO-PRE	ICS 89 / ELI 36 / ELI 37 / EMV 52
17.605	DMU	ELMO-STA	SPS 1000 / ELV 49 / ELV 68 / EMV 26
26.600	DMU	ELMO-STA	DSK611 / ELV 49 / ELV 68 / EMV 26

[6] **Test Report**

The proof of explosion protection of the equipment stated under [2] b) is documented in the Test Report IB-11-3-118 of 2 February 2012. The test documents are part of the Test Report.


[7] **Test result**

IBExU certifies that the equipment stated under [2] b) fulfils the Essential Health and Safety Requirements given in Annex II to the Directive 94/9/EC by compliance with EN 60079-0:2009, EN 60079-11:2007 and EN 61241-11:2006.

The equipment stated under [2] b) fulfils the requirements of type of protection Intrinsic Safety 'ia' for an electrical equipment of the Equipment Group II, Category 1G, Explosion Group IIC and Temperature Class T4 as well as Category 1D with a maximum surface temperature of 85 °C.

The marking of the equipment stated under [2] b) shall include the following:

 II 1G Ex ia IIC T4

 II 1D Ex iaD 20 T 85 °C

**IBExU Institut für Sicherheitstechnik GmbH**  
An-Institut der TU Bergakademie Freiberg

This addition is only valid in combination with the EC-Type Examination Certificate  
IBExU10ATEX1122 X of 31 August 2010 and their 1<sup>st</sup> Addition.

**IBExU** Institut für Sicherheitstechnik GmbH  
Fuchsmühlenweg 7 - 09599 Freiberg, Germany  
☎ +49 (0) 3731 3805-0 - ☎ +49 (0) 3731 23650

Authorized for certifications  
- Explosion protection -

By order



(Dr. Wagner)



- Seal -  
(Identification No. 0637)

Freiberg, 2 February 2012

Certificates without signature and seal are not valid.  
Certificates may only be duplicated completely and unchanged.  
In case of dispute, the German text shall prevail.

- [1] **3<sup>rd</sup> Addition to**  
**EC-TYPE EXAMINATION CERTIFICATE IBExU10ATEX1122 X**  
according to Directive 94/9/EC Annex III  
**- Translation -**



- [2] Equipment: **Pressure transmitter**
- a) Types corresponding IBExU10ATEX1122 X:  
DX9-DMP..., DX9-DMK..., DX9-LMP...,  
DX9-LMK...and DX9-17.600G
- b) Added/appended types: DX9-DMP 334, DX9-DMP 311 and  
DX9-DMP 321
- [3] Manufacturer: BD SENSORS s.r.o.
- [4] Address: Hradistska 817  
687 08 Buchlovice  
CZECH REPUBLIC

[5] **Addition / Alteration**

The additions / alterations of the equipment stated under [2] compared to the equipment already certified comprise in particular:

- a) Addition of types DX9-DMP 334, DX9-DMP 311 and DX9-DMP 321 under use of new measuring cells as well as electronic configurations on the basis of the partial certification IBExU10ATEX1026 U incl. its Additions.

The table is added as follows:

Name	Type	Electronic-module	Measuring cell
DMP 334	DMU	ELMO-STA ELV68 with EMV26 or ELV93 with EMV65	ST1 with ELV97
DMP 311	DMU	ELMO-STA ELV68 with EMV26	DSP4XX. ICS86, NPI15 with ELV66,ELV70, ELV86
DMP 321	DMU	ELMO-PRE ELI37 with EMV 52	DSP4XX, ICS86, NPI15, ST1 with ELV79, ELV67, ELI68, ELI69, ELI57, ELI63, ELI73
DMK 331*	DMU	ELMO-STA ELV68 with EMV26 or ELV93 with EMV65	DSK 516 with ELV 94

\* Type with a different configuration already exists

- b) The devices can also be produced with the amplifier electronics ELV9 with EMV26 and the measuring cells without preamplifier. When using the electronics ELV9 the same safety limit values shall be as in the previous ELMO-STA-modules.
- c) For the devices DX9-DMP331 and DX9-DMP333 with fixed-laid cable the use was extended to a lower temperature range to -40 ° C.
- d) The devices also fulfils the requirements of the current standards of EN 60079-0:2012.

[6] **Test Report**

The proof of the explosion protection of the equipment stated under [2] is documented in the Test Report IB-13-3-256 of 22 January 2014. The test documents are part of the Test Report.



[7] **Test result**

IBExU certifies that the equipment stated under [2] fulfils the Essential Health and Safety Requirements given in Annex II to the Directive 94/9/EC by compliance with EN 60079-0:2012, EN 60079-11:2012 and EN 60079-26:2007.

The equipment stated under [2] fulfils the requirements of type of protection Intrinsic Safety 'ia' for an electrical equipment of the Equipment Group II, Category 1G, Explosion Group IIC and Temperature Class T4 as well as Category 1D with a maximum surface temperature of 85 °C.

The marking of the equipment stated under [2] shall include the following:

Non-metallic pressure connection	⊕ II 2G Ex ia IIC T4 Gb
	⊕ II 2D Ex ia IIIC T 85 °C Db
Metallic pressure connection	⊕ II 1G Ex ia IIC T4 Ga
	⊕ II 1D Ex ia IIIC T 85 °C Da

[8] **Special conditions**

The special conditions mentioned in the EC-Type Examination Certificate have to be taken into account furthermore.

**This addition is only valid in combination with the EC-Type Examination Certificate IBExU10ATEX1122 X of 31 August 2010 and its Additions.**


**IBExU** Institut für Sicherheitstechnik GmbH  
Fuchsmühlenweg 7 - 09599 Freiberg, Germany  
☎ +49 (0) 3731 3805-0 - 📠 +49 (0) 3731 23650

Authorized for certifications  
- Explosion protection -

By order



(Dr. Wagner)



- Seal -  
(Identification No. 0637)

Freiberg, 22 January 2014

Certificates without signature and seal are not valid.  
Certificates may only be duplicated completely and unchanged.  
In case of dispute, the German text shall prevail.