

EU DECLARATION OF CONFORMITY

Act No. 90/2016 Coll., Act No. 265/2017 Coll. and Act No. 91/2016 Coll. as amended

BD SENSORS s.r.o., Hradištská 817, 687 08 Buchlovice, Czech Republic, ID-Nr 49968416

declares on its own responsibility that the


products: precise digital pressure gauges**types:** AX6-DM01, AX6-DM01-500, AX6DM01-500 HD, AX6-DL102 with DCL552M

match the under mentioned requirements:

Harmonized standards of documents	
ČSN EN IEC 60079-0 ed. 5:2018	EN IEC 60079-0:2018
ČSN EN 60079-11 ed. 2:2012	EN 60079-11:2012
ČSN EN 60079-26 ed. 3:2015	EN 60079-26:2015

Czech and European directives		
NV 116/2016 Sb	2014/34/EU	(Ex Directive)
NV 117/2016 Sb.	2014/30/EU	(EMC Directive)
NV 391/2016 Sb.	2015/863/EU	(RoHs Directive)

The marking of the products must include one of the following:

-  **II 1G Ex ia IIC T4 Ga** (variant with conductive foil)
- II 2G Ex ia IIB T4 Gb** (standard variant)
- II 2G Ex ia IIB T4 Gb** (AX6-DL102 with AX6-DCL552 M)

Basis for this EU Declaration of Conformity are reports and documents from following organizations:

- *IBExU Institut für Sicherheitstechnik GmbH*
Fuchsmühlweg 7,
D-09599 Freiberg

EC-Type Examination Certificate Number IBExU12ATEX1107 X from 28.08.2012 (test report IB-11-3-259) Issue 1 of this certificate from 10.07.2017 (test report IB-17-3-0079) and Issue 2 of this certificate from 01.10.2018 (test report IB-18-3-0093).

- *FTZÚ Ostrava-Radvanice, notified body number 1026*
Quality Assurance Notification FTZÚ 13 ATEX Q 007 from 13.07.2022
(ČSN EN ISO/ITEC 80079-34 ed.2 : 2020, NV 116/2016 Sb.)

Buchlovice 02.05.2023

Ing. Martin Šimoník
Head of Electro Development DepartmentIng. Radek Burda
Ex Appointee

EU PROHLÁŠENÍ O SHODĚ

dle zákona č. 90/2016 Sb., zákona č. 265/2017 Sb. a zákona č. 91/2016 Sb. v platném znění

My, BD SENSORS s.r.o., Hradištská 817, 687 08 Buchlovice, Česká republika, IČO 49968416

prohlašujeme na svou výlučnou odpovědnost, že

výrobky: precizní digitální manometry


typy: AX6-DM01, AX6-DM01-500, AX6DM01-500 HD, AX6-DL102 s DCL552M

jsou ve shodě s následujícími normami:

České normy	Evropské normy
ČSN EN IEC 60079-0 ed. 5:2018	EN IEC 60079-0:2018
ČSN EN 60079-11 ed. 2:2012	EN 60079-11:2012
ČSN EN 60079-26 ed. 3:2015	EN 60079-26:2015

a následujícími nařízeními vlády (NV), v závorce uvedeno číslo EU směrnice:

NV 116/2016 Sb. v platném znění	2014/34/EU v platném znění	(Ex)
NV 117/2016 Sb. v platném znění	2014/30/EU v platném znění	(EMC)
NV 391/2016 Sb. v platném znění	2015/863/EU v platném znění	(RoHs)

Výrobek, za podmínek uvedených v návodu k použití a v ES Certifikátu o přezkoušení typu, splňuje stupeň nevybušnosti  **II 1G Ex ia IIC T4 Ga** (varianta s vodivou čelní fólií)

II 2G Ex ia IIB T4 Gb (standardní varianta)

II 2G Ex ia IIB T4 Gb (AX6-DL102 s AX6-DCL552 M)

- *IBExU Institut für Sicherheitstechnik GmbH*

*Fuchsmühlweg 7,
D-09599 Freiberg*

Zkoušel dané výrobky a vydal:

ES certifikát o přezkoušení typu č. IBExU12ATEX1107 X z 28.08.2012 (zkušební zpráva IB-11-3-259)

Vydání 1 tohoto certifikátu z 10.07.2017 (zkušební zpráva IB-17-3-0079) a Vydání 2 tohoto certifikátu z 01.10.2018 (zkušební zpráva IB-18-3-0093).

FTZÚ Ostrava-Radvanice, oznámený subjekt č. 1026 vydal:

Oznámení o zabezpečování kvality č. FTZÚ 13 ATEX Q 007 z 13.07.2022

(ČSN EN ISO/ITEC 80079-34 ed.2 : 2020, NV č. 116/2016 Sb.)

Buchlovice 02.05.2023

Ing. Martin Šimoník
Vedoucí vývoje elektro

Ing. Radek Burda
Pověřená osoba Ex

